

# Cyberknife treatment plan MU and Dose verification using IMSureQA





Anthony Lo

Stanford University Medical Center, Stanford CA, USA

## Introduction

Cyberknife Treatment plans generated from MultiPlan 2.1.0 system are checked by commercial software IMSureQA 3.1.0 (Standard Imaging, Inc.) MultiPlan uses Ray-Tracing as the default dose calculation mode with beam parameters: OF (output factor), TPR (tissue phantom ratio) and OCR (off center ratio). These beam parameters are loaded into IMSureQA through the Physics module. The IMSureQA program checks total dose to a reference point defined during planning. Beam parameters and dose contribution to the reference point from individual beam are stored in Beam List and total reference dose is stored in Plan Overview. IMSureQA calculates dose from individual beam and sum as the total dose, which is compared with the reference dose from Plan Overview. Over 200 plans have been checked using IMSureQA for 2 Cyberknives since Feb 2008. Results from recent plans are summarized.

## Beam Data



## Sample IMSureQA Result

## Variation Summary

|             | Total dose difference % | % of beams >3% | % of beams >5% |                      |
|-------------|-------------------------|----------------|----------------|----------------------|
| 6D only     | 0.2<br>(0.2)            | 7<br>(4)       | 2<br>(4)       | average<br>(std dev) |
| Xsight only | 1.0<br>(0.2)            | 9<br>(3)       | 4<br>(3)       | Average<br>(std dev) |
| Overall     | 0.4<br>(0.4)            | 7<br>(4)       | 3<br>(4)       | average<br>(std dev) |

## Results

Average deviation of total dose between MultiPlan and IMSureQA is 0.4%. 7% of all beams are > 3% dose difference and 3% of all beams are > 5% dose difference. Xsight tracking plans average 1% total dose deviation, 9% beams > 3% dose difference and 4% beams > 5% dose difference. Beam with large dose difference is due to dose calculation point at large off axis distance and/or far away from calibration distance of 80 cm. Simple isocentric plans have the smallest deviation while complex plans for irregular shaped target or multiple targets had the biggest deviation.

## Variation Between Multiplan and IMSureQA

| Case      | Cone     | Total dose difference % | >3%  | >5%  | Target size and shape |
|-----------|----------|-------------------------|------|------|-----------------------|
| iso-1     | 7.5      | 0                       | 0    | 0    |                       |
| iso-2     | 12.5     | 0                       | 0    | 0    |                       |
| iso-3     | 15       | 0.1                     | 0    | 0    |                       |
| 6D-1      | 10       | -0.2                    | 2.9  | 0    |                       |
| 6D-2      | 16       | -0.1                    | 5.8  | 0.7  | large long long       |
| 6D-3      | 10       | -0.1                    | 14.7 | 14.7 |                       |
| 6D-4      | 15       | -0.1                    | 4.9  | 2.7  |                       |
| 6D-5      | 15       | -0.1                    | 15.3 | 14.7 | long                  |
| 6D-6      | 10       | 0                       | 3    | 0.6  |                       |
| 6D-7      | 20       | 0.1                     | 7.4  | 4    |                       |
| 6D-8      | 7.5      | 0.1                     | 8.7  | 0    |                       |
| 6D-9      | 12.5     | 0.2                     | 5.8  | 0    |                       |
| 6D-10     | 7.5      | 0.2                     | 6.7  | 0    |                       |
| 6D-11     | 10       | 0.2                     | 6.7  | 0    |                       |
| 6D-12     | 15       | 0.2                     | 7.3  | 0.7  |                       |
| 6D-13     | 7.5      | 0.3                     | 8.8  | 1.5  |                       |
| 6D-14     | 12.5     | 0.3                     | 4.4  | 0    |                       |
| 6D-15     | 10       | 0.3                     | 9    | 1.2  |                       |
| 6D-16     | 10       | 0.3                     | 2.7  | 0    |                       |
| 6D-17     | 10       | 0.4                     | 6    | 1.2  |                       |
| 6D-18     | 15       | 0.4                     | 8.8  | 1.9  |                       |
| 6D-19     | 5        | 0.4                     | 11.3 | 0.7  |                       |
| 6D-20     | 5        | 0.4                     | 3.3  | 0    |                       |
| Xsight-1  | 12.5     | 0.4                     | 10.2 | 4.1  | large long            |
| Xsight-2  | 10       | 0.7                     | 12   | 9.3  |                       |
| Xsight-3  | 7.5      | 1                       | 9.8  | 0.8  | irregular             |
| Xsight-4  | 12.5, 20 | 1                       | 9.9  | 6.8  |                       |
| Xsight-5  | 25       | 1                       | 12.8 | 5    | irregular long        |
| Xsight-6  | 15       | 1.1                     | 6.7  | 1.9  |                       |
| Xsight-7  | 20       | 1.1                     | 6.5  | 4.3  | long                  |
| Xsight-8  | 20       | 1.1                     | 9.4  | 6    |                       |
| Xsight-9  | 15, 35   | 1.2                     | 6.7  | 2.4  | irregular             |
| Xsight-10 | 25       | 1.2                     | 4.7  | 0    |                       |

## Conclusions

IMSureQA consistently calculates MU / dose with under 1% deviation from MultiPlan. Large individual beam dose variation comes from large off axis distance which contributes very little dose to the reference point therefore the total dose deviation is quite small. IMSureQA can replace hand calculation for MU dose verification purpose.

